

Suggested Cleaning Recommendation

Nonin Medical, Inc. recommends the following process for cleaning your device:

Wipe the device with a soft cloth dampened with a mild detergent or 10% bleach solution. Do not use undiluted bleach or any cleaning solution other than those recommended here, as permanent damage could result. Dry with a soft cloth, or allow to air dry.

Caution: Do not place the device in liquid or clean it with agents containing ammonium chloride, isopropyl alcohol, or products that are not listed in the User's Guide.